

**SYNERGY**<sup>™</sup>

Everolimus-Eluting Platinum Chromium Coronary Stent System

IN-SERVICE PRESENTATION


WITH CONFIDENCE

#### **Vascular Response to Current Permanent Polymers**

The case for bioabsorbable polymer coatings



#### Focal Inflammation




CoCr EES
Focal inflammation
with eosinophils
(4 months)

### Chronic Inflammation




CoNi ZES
Chronic inflammation
with giant cells
secondary to polymer
delamination
(3 months)

#### Neo-Atherosclerosis



CoCr EES
Foamy macrophage
accumulation
(neoatherosclerosis)


### Late Stent Thrombosis



CoCr EES
EES implanted
within PES
(6 months
antemortem)


#### What is the purpose of a DES polymer?

- Provide a mechanically stable matrix for the drug
- Modulate drug release into the vessel wall



#### What happens after drug release?

- Polymer remaining after drug release has no function
- All polymer coatings have the potential to be damaged
- Permanent polymers are permanent




#### **Clinical Implications**


- Late / very late events
- Chronic inflammation with neoathersclerosis
- Constant irritant may lead to late restenosis

#### Bioabsorbable Polymer in Perspective Relative Polymer and Drug Absorption Profiles



#### The SYNERGY Stent's polymer is absorbed shortly after drug elution ends at 3 months





#### SYNERGY™ Stent

Best-in-Class Visibility



<sup>\*</sup>Testing Completed by Boston Scientific data on file. 2.5 mm stent products tested. Based on 2.5mm stents. Under 6.0mm copper phantom to simulate body mass. Bench test results may not necessarily be indicative of clinical performance.

<sup>\*\*</sup>Strut thickness for small vessel model is 74µm, Workhorse model is 79µm and large vessel is 81µm. Boston Scientific data on file.

SYNERGY<sup>1</sup> Stent

# Thin Struts and Lower Coating Thickness Xience V<sup>™</sup>, Xience Prime<sup>™1</sup>, Xience Xpedition<sup>™1</sup>, Xience Alpine<sup>™</sup> Stents PROMUS Element<sup>™1</sup> Stent Promus PREMIER<sup>™</sup> Stent

Strut Thickness	74 μm* (0.0029")	81 μm (0.0032")	120 μm (0.0047")	150 μm (0.0059")
Polymer Coating Type & Thickness	Bioabsorbable Abluminal 4 μm	Conformal Permanent 8 µm	Bioabsorbable Abluminal 10 µm	Bioabsorbable Conformal 3 µm / side
Total Coated Strut Thickness	78 µm	97 µm	130 µm	156 µm

\*Strut thickness for small vessel model is 74µm, Workhorse model is 79µm and large vessel is 81µm. Boston Scientific data on file. Representative drawings are to scale. Representative drawings. 1 Wilson GW, EuroIntervention 2012;8:250-257, and Soucy NV, EuroIntervention 2010;6:630-637. 2 Gutiérrez-Chico JL, American Heart Journal November 2011; 162:922–931. 3 Onuma Y, Circulation 2011, 123:779-797. Bench test results may not necessarily be indicative of clinical performance.

#### SYNERGY™ Stent

#### 1 month DAPT labelling update now CE marked



#### **Individualisation of Patient Treatment**

The device carries an associated risk of acute, sub-acute or late thrombosis, vascular complications, and/or bleeding events. Therefore, patients should be carefully selected, and a P2Y<sub>12</sub> inhibitor (i.e., clopidogrel, ticlopidine, prasugrel, or ticagrelor) must be prescribed post procedure to reduce the risk of stent thrombosis. Aspirin must be administered concomitantly with the P2Y<sub>12</sub> inhibitor, and then continued indefinitely to further reduce the risk of thrombosis.

Antiplatelet drugs should be used in combination with SYNERGY which is designed with a low initial polymer load, abluminal coating and bioabsorbable polymer which may reduce the risk of thrombosis and the need for prolonger dual antiplatelet therapy.



Physicians should use the information from the large body of clinical evidence for everolimus drug eluting stents, coupled with current literature on drug-eluting stents, current European Society of cardiology recommendation or other applicable country guidelines and the specific needs of the individual patient to determine the specific antiplatelet / anticoagulation regimen to be used for their patients in general practice.

It is very important that the patient be compliant with post-procedural antiplatelet recommendations given by their physician. In selected higher risk patients where the physician determines that the risks outweigh the benefits of continued DAPT, it may be reasonable to interrupt or discontinue therapy after 1 month based on low stent thrombosis rates and no observed increased risk for stent thrombosis as shown in the current literature. Patients who require premature discontinuation of antiplatelet therapy should be monitored closely and have their antiplatelet therapy restarted as soon as possible per the discretion of their treating physician.

<sup>^</sup> The updated Directions for Use will also include an update on the current "Pre-and Post-Procedure Antiplatelet Regimen":

In selected higher risk patients where the physician determines that the risks outweigh the benefits of continued DAPT, it may be reasonable to interrupt or discontinue therapy after 1 month based on low stent thrombosis rates and no observed increased risk for stent thrombosis as shown in the current literature. Patients who require premature discontinuation of antiplatelet therapy should be monitored closely and have their antiplatelet therapy restarted as soon as possible per the discretion of their treating physicians

#### SYNERGY™ Coronary Stent System

New Indications and Labelling Updates Now CE Marked



#### **CE Mark**

- ✓ Acute Coronary Syndrome (ACS)
- ✓ Acute Myocardial Infarction (AMI)
- ✓ Unstable Angina
- ✓ Renal Failure
- ✓ Coronary Bifurcation Lesions¹
- ✓ Coronary Multi-vessel Disease
- ✓ Coronary Saphenous Vein Graft Lesions
- ✓ Coronary Artery Ostial Lesions
- ✓ Unprotected Left Main Coronary Artery Lesions
- ✓ Coronary Artery Total Occlusion Lesions<sup>2</sup>
- ✓ In-Stent Restenosis in Coronary Artery Lesions³

<sup>&</sup>lt;sup>1</sup> When treating Bifurcations, care must be exercised to access the secondary vessel via the repeating geometry in the body of the stent within the primary vessel.


<sup>&</sup>lt;sup>2</sup> For treatment of occluded vessels, contrast visualization of the distal vessel to confirm position of guidewire within the lumen is recommended.

<sup>&</sup>lt;sup>3</sup> For in stent restenosis, where details of the original stent are known, the expanded inner diameter of the new stent should not exceed the dilation limits of the original stent. Where details of the original stent are not known, the expanded inner diameter of the new stent should not exceed the reference vessel diameter.



- Left Main vessels are Large (>5.5mm on average¹)
- SYNERGY™ Stent is indicated for treatment of patients presenting with unprotected left main coronary artery lesions
- SYNERGY has a labeled overexpansion indication of up to 5.75mm<sup>2</sup>
- Left Main Vessels require stents with significant radial strength<sup>3</sup>
- SYNERGY offers excellent radial strength<sup>4</sup>
- Left Main PCI needs optimal Vessel Healing<sup>3</sup>
- The SYNERGY DES shows healing within 3-month and has low risk of cardiac adverse events in complex patients<sup>5</sup>

**IVUS LM in male patients** 



Graph Prof. Robert Jan van Geuns. IVUS LM in male pts (Rotterdam). Asia PCR 2015.

IVUS data on 865 LAD->LM pullbacks, male patients . Average Vessel Diameter (VD):  $5.60\,\mathrm{mm}$  , Mean VD:  $5.47\,\mathrm{mm}$ . Only in 13% <5 mm.

<sup>1.</sup> Shand J, et al. Prospective Intravascular Ultrasound Investigation of the Necessity for and Efficacy of Postdilation Beyond Nominal Diameter of 3 Current Generation DES Platforms for the Percutaneous Treatment of the Left Main Coronary Artery. Cathet Cardiovasc Intv; 2014;84:351-358.

<sup>2.</sup> Labeled Post-Dilatation Limits. SYNERGY Stent, Xience Xpedition Stent , Resolute Integrity Stent and Resolute Onyx DFU

<sup>3.</sup> Adapted from presentations by Prof. RJ van Geuns, MD, PhD; and Jiang Ming Fam, MD at AsiaPCR 2015

<sup>4.</sup> Bench testing performed by Boston Scientific Corporation. Data on file at Boston Scientific. All stents 2.5mm. SYNERGY Stent n=5, all other stents are n=3. Bench test results not necessarily indicative of clinical performance.

<sup>5.</sup> EVOLVE Clinical trial at 4 year. Presented by Ian T. Meredith AM, MBBS, PhD, PCR 2015. EVOLVE II Clinical Trial. Presented by Dean J. Kereiakes, MD at AHA 2014. EVOLVE II DM Clinical Trial. Presented by Stephan Windecker. MD at PCR 2015. Adapted from a presentation by J. M. de la Torre. MD at PCR 2015. Wilson G. MD. ACC 2011.

#### SYNERGY™ Stenting Total Occlusion


Scientific

Ac e for life

CTO Chronic Total Occlusion

- SYNERGY™ Stent is indicated for treatment of patients presenting with total occlusion coronary artery lesions:
- CTO lesions tend to be long and located in large vessels
- SYNERGY has a labeled overexpansion of up to 5.75mm, allowing the physician to customize the stent to the appropriate vessel size<sup>1</sup>
- Stent mechanical properties are important considerations for CTO stenting
- The SYNERGY Stent's customized architecture offers exceptional strength and conformability<sup>2</sup>
- The SYNERGY DES shows healing within 3-months and has low risk of cardiac adverse events in complex patients<sup>3</sup>





For treatment of occluded vessels with the SYNERGY stent system, contrast visualization of the distal vessel to confirm position of guidewire within the lumen is recommended.

1 Labeled Post-Dilatation Limits. SYNERGY Stent, Xience Xpedition Stent, Resolute Integrity Stent, and Resolute Onyx DFU

3 EVOLVE Clinical trial at 4 year. Presented by Ian T. Meredith AM, MBBS, PhD, PCR 2015. EVOLVE II Clinical Trial. Presented by Dean J. Kereiakes, MD at AHA 2014. EVOLVE II DM Clinical Trial. Presented by Stephan Windecker, MD at PCR 2015. Adapted from a presentation by J. M. de la Torre, MD at PCR 2015. Wilson G, MD. ACC 2011. These studies did exclude CTO lesions.

<sup>2</sup> Bench testing performed by Boston Scientific Corporation. Data on file at Boston Scientific. All stents 2.5mm. SYNERGY Stent n=5, all other stents are n=3. Bench test results not necessarily indicative of clinical performance.

#### SYNERGY™ Stenting Bifurcation

# Scientific Agree for life\*\* BIF Bifurcation Lesions

# SYNERGY™ Stent is indicated for treatment of patients presenting with bifurcation coronary artery disease:

- Bifurcation lesions are often tapered<sup>1</sup>
- SYNERGY has a labeled overexpansion of up to 5.75mm, allowing the physician to customize the stent to the appropriate vessel size and ensure great apposition<sup>2</sup>
- With bifurcation we want to maintain the natural vessel shape<sup>1</sup>
- The SYNERGY Stent's customized architecture offers exceptional strength and conformability<sup>3</sup>
- Bifurcation PCI benefits from an appropriate Cell Diameter & Expansion<sup>1</sup>
- The SYNERGY Stent has large cell diameters in the body of the stent to accommodate side branch access<sup>4</sup>
- Bifurcation lesions need optimal Vessel Healing to reduce risk of ST<sup>1</sup>
- The SYNERGY DES shows healing within 3-month and has low risk of cardiac adverse events in complex patients<sup>5</sup>



When treating Bifurcations, care must be exercised to access the secondary vessel via the repeating geometry in the body of the stent within the primary vessel.

1 Attributes collected from: Curtiss T. Stinis, M.D., F.A.C.C, F.S.C.A.I. - Scripps CI 2013 and ESC/EACTS GUIDELINES 2014. EHJ doi:10.1093/eurheartj/ehu278. 2 Labeled Post-Dilatation Limits. SYNERGY Stent, Xience Xpedition Stent, Resolute Integrity Stent and Resolute Onyx DFU 3 Bench testing performed by Boston Scientific Corporation. Data on file at Boston Scientific. All stents 2.5mm. SYNERGY Stent n=5, all other stents are n=3. Bench test results not necessarily indicative of clinical performance. 4 Circular Cell Diameter: n = 5. Data on file at Boston Scientific. Size Matrix on the SYNERGY Stent DFUs. 5 EVOLVE Clinical trial at 4 year. Presented by Ian T. Meredith AM, MBBS, PhD, PCR 2015. EVOLVE II Clinical Trial. Presented by Dean J. Kereiakes, MD at AHA 2014. EVOLVE II DM Clinical Trial. Presented by Stephan Windecker, MD at PCR 2015. Adapted from a presentation by J. M. de Ia Torre, MD at PCR 2015. Wilson G, MD. ACC 2011.

#### SYNERGY™ Stenting Bifurcation

# Scientific Address to the second sec

# SYNERGY™ Stent is indicated for treatment of patients presenting with bifurcation coronary artery disease:

- Bifurcation lesions are often tapered<sup>1</sup>
- SYNERGY has a labeled overexpansion of up to 5.75mm, allowing the physician to customize the stent to the appropriate vessel size and ensure great apposition<sup>2</sup>
- With bifurcation we want to maintain the natural vessel shape<sup>1</sup>
- The SYNERGY Stent's customized architecture offers exceptional strength and conformability<sup>3</sup>
- Bifurcation PCI benefits from an appropriate Cell Diameter & Expansion<sup>1</sup>
- The SYNERGY Stent has large cell diameters in the body of the stent to accommodate side branch access<sup>4</sup>
- Bifurcation lesions need optimal Vessel Healing to reduce risk of ST<sup>1</sup>
- The SYNERGY DES shows **healing within 3-month** and has low risk of cardiac adverse events **in complex patients**<sup>5</sup>




When treating Bifurcations, care must be exercised to access the secondary vessel via the repeating geometry in the body of the stent within the primary vessel.

1 Attributes collected from: Curtiss T. Stinis, M.D., F.A.C.C, F.S.C.A.I. - Scripps CI 2013 and ESC/EACTS GUIDELINES 2014. EHJ doi:10.1093/eurheartj/ehu278. 2 Labeled Post-Dilatation Limits. SYNERGY Stent, Xience Xpedition Stent, Resolute Integrity Stent and Resolute Onyx DFU 3 Bench testing performed by Boston Scientific Corporation. Data on file at Boston Scientific. All stents 2.5mm. SYNERGY Stent n=5, all other stents are n=3. Bench test results not necessarily indicative of clinical performance. 4 Circular Cell Diameter: n = 5. Data on file at Boston Scientific. Size Matrix on the SYNERGY Stent DFUs. 5 EVOLVE Clinical trial at 4 year. Presented by Ian T. Meredith AM, MBBS, PhD, PCR 2015. EVOLVE II Clinical Trial. Presented by Stephan Windecker, MD at PCR 2015. Adapted from a presentation by J. M. de Ia Torre, MD at PCR 2015. Wilson G, MD. ACC 2011.



#### **NEW Laser-Cut Hypotube**

- ~300 cuts over
   100 mm length
- Extends into midshaft to exit port to improve pushability
- Additional length maintains midshaft flexibility
- 个 Pushability and Flexibility




#### SYNERGY™ Stent System

SYNERGY and PREMIER 4.0mm dedicated stent, with 1.75mm overexpansion capability, is more versatile when sizing for the

LAD and LCX and over-expanding to fit the LM

Labeled Post-Dilatation Limits\*






#### Kang Network Meta-Analysis: Relative ST risk at 1-year from 111,088 patients



SYNERGY ™ BP Stent Ranked #1 and Promus PREMIER ™ PP Stent Ranked #2 for the lowest relative risk of Definite/Probably Stent Thrombosis

#### **Promus PREMIER vs. comparators**



#### **STUDY DESIGN (ST Analysis)**


- 110 prospective, randomized controlled trials included
- 111,088 patients
- Primary endpoint: definite or probable stent thrombosis at 1 year

# SYNERGY™ Stent System Summary



## Providing operational efficiency, optimal<sup>1</sup> healing, and freedom from long-term polymer exposure

- SYNERGY™ Stent offers peace of mind knowing that the polymer is gone shortly after the drug is completely eluted at 3 months²
- SYNERGY Stent has shown full stent strut coverage as early as 30-days in humans and showed 0% stent thrombosis in multiple OCT analyses.<sup>3</sup>
- Unmatched acute performance which reduces procedure time as much as 5 minutes per case.<sup>4</sup>



<sup>&</sup>lt;sup>1</sup> Eppihimer M, PhD. Impact of Polymer Type and Location on Stent Thrombogenicity and Endothelial Cell Coverage. EuroPCR 2014.

<sup>&</sup>lt;sup>2</sup> Wilson, G.J., et al. Catheter Cardiovasc Interv. 2015.

<sup>&</sup>lt;sup>3</sup> OCT Analyses: SORT OUT VIII - Adapted from presentation by Ida Riise Balleby, MD at PCR 2015; OCT Analysis presented by J. M. de la Torre, MD at TCT 2014; TIMELESS - Adapted from presentation by Juan Granada, MD, at CRT 2015; Burgos-Santander Study - Adapted from a presentation by J. M. de la Torre, MD at PCR 2015.

<sup>&</sup>lt;sup>4</sup> US IC Blinded Product Perception survey (62 ICs), Increased stent visibility was estimated to decrease stent usage by 14% balloon usage by 18% during a typical PCI and 5 minutes less procedure time in IC Perception survey; 95% (59/62) identified the PtCr stent as the most visible.